

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 11/22/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:41
INTERNAL STANDARD	VERIFIED	04:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	04:43
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:44
BLANK TEST	0.00	04:45

