

SUPERVISOR TEST

-----  
HUDSON FALLS  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 102408  
-----

Date: 12/22/2020

OPERATOR NAME:  
BRAD M LYON  
Permit #: 44978  
Operator Agency/Dept: HUDSON FALLS VIL. PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:07
INTERNAL STANDARD	VERIFIED	23:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:07
BLANK TEST	0.00	23:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:09
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:11

