

COLLECTOR 1234

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 12/20/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:24
INTERNAL STANDARD	VERIFIED	07:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:25
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:26
BLANK TEST	0.00	07:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28

