

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 10/17/2020

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:55
INTERNAL STANDARD	VERIFIED	09:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:58
BLANK TEST	0.00	09:59

