

WHITEHALL VILLAGE
 POLICE DEPT.
 DATAMASTER dmt: 113806

Date: 10/10/2020

OPERATOR NAME:
 BRYAN J GRECO
 Permit #: 44077
 Operator Agency/Dept: WHITEHALL VILL. P.D.
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:00
INTERNAL STANDARD	VERIFIED	09:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:02
BLANK TEST	0.00	09:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04

