

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 12/26/2020

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:46
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:48
BLANK TEST	0.00	22:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:49
BLANK TEST	0.00	22:50

