

SOLUTION CHANGE

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 11/02/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:22
INTERNAL STANDARD	VERIFIED	06:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:25
BLANK TEST	0.00	06:26

