

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 07/27/2020

OPERATOR NAME:  
CARLEIGH B BATCHELDER  
Permit #: 48533  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:12
INTERNAL STANDARD	VERIFIED	14:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:14
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:16
BLANK TEST	0.00	14:17

CB #708

