

SOLUTION CHANGE

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 08/03/2020

OPERATOR NAME:
 CORY L HURLBURT
 Permit #: 37017
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:21
INTERNAL STANDARD	VERIFIED	06:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:22
BLANK TEST	0.00	06:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:25
BLANK TEST	0.00	06:26

Handwritten: # 257

