

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 08/24/2020

OPERATOR NAME:  
KYLE L KOLAR  
Permit #: 44063  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:24
INTERNAL STANDARD	VERIFIED	22:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:26
BLANK TEST	0.00	22:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:27
BLANK TEST	0.00	22:28

