

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 09/08/2020

OPERATOR NAME:  
CARLEIGH B BATCHELDER  
Permit #: 48533  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:00
INTERNAL STANDARD	VERIFIED	17:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:01
BLANK TEST	0.00	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:02
BLANK TEST	0.00	17:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04

*Batchelder*  
*#708*

