

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/14/2020

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:56
INTERNAL STANDARD	VERIFIED	06:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:57
BLANK TEST	0.00	06:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:58
BLANK TEST	0.00	06:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:59
BLANK TEST	0.00	07:00

#821
CM

