

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 09/29/2020

OPERATOR NAME:  
WILLIAM M GREENE  
Permit #: 48760  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:09
BLANK TEST	0.00	00:10

