

SOLUTION CHANGE

WASHINGTON COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 106406

Date: 10/05/2020

OPERATOR NAME:  
 CORY L HURLBURT  
 Permit #: 37017  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:25
INTERNAL STANDARD	VERIFIED	07:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:26
BLANK TEST	0.00	07:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:28
BLANK TEST	0.00	07:29

CH 257

