

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/12/2020

OPERATOR NAME:
CORBET W SULLIVAN
Permit #: 45898
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:52
INTERNAL STANDARD	VERIFIED	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:52
BLANK TEST	0.00	06:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56

CS
#623

