

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/26/2020

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:16
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20

CM
#821

