

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 11/16/2020

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 4876020
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:11
INTERNAL STANDARD	VERIFIED	19:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:13
BLANK TEST	0.00	19:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:14
BLANK TEST	0.00	19:15



#647

