

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 11/30/2020

OPERATOR NAME:
 ETHAN J MACURA
 Permit #: 44080
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:15
INTERNAL STANDARD	VERIFIED	14:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:17
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:18
BLANK TEST	0.00	14:19

EJ Macura
 #694

