

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/09/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:31
INTERNAL STANDARD	VERIFIED	13:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:32
BLANK TEST	0.00	13:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:33
BLANK TEST	0.00	13:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:34
BLANK TEST	0.00	13:35

