

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 11/16/2020

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:52
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:53
BLANK TEST	0.00	05:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	05:55
BLANK TEST	0.00	05:56

