

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 11/30/2020

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:57
INTERNAL STANDARD	VERIFIED	21:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:00
BLANK TEST	0.00	22:00

