

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/01/2020

OPERATOR NAME:
ROBERT SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:35
INTERNAL STANDARD	VERIFIED	01:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:36
BLANK TEST	0.00	01:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:37
BLANK TEST	0.00	01:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:38
BLANK TEST	0.00	01:39

