

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 12/14/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:07
INTERNAL STANDARD	VERIFIED	17:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:08
BLANK TEST	0.00	17:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11

