

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 08/17/2020

OPERATOR NAME:
 MATTHEW K JACKSON
 Permit #: 40078
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:14
INTERNAL STANDARD	VERIFIED	14:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:15
BLANK TEST	0.00	14:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:16
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:17
BLANK TEST	0.00	14:18

