

SUPERVISOR TEST

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 09/14/2020

OPERATOR NAME:  
MATTHEW K JACKSON  
Permit #: 40078  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:02
INTERNAL STANDARD	VERIFIED	17:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:03
BLANK TEST	0.00	17:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:04
BLANK TEST	0.00	17:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:05
BLANK TEST	0.00	17:06

