

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/28/2020

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:21
INTERNAL STANDARD	VERIFIED	10:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:21
BLANK TEST	0.00	10:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:22
BLANK TEST	0.00	10:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:23
BLANK TEST	0.00	10:24

