

SUPERVISOR TEST

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 08/19/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:23
INTERNAL STANDARD	VERIFIED	00:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:24
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27

