

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 11/25/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:09
INTERNAL STANDARD	VERIFIED	19:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:11
BLANK TEST	0.00	19:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	19:12
BLANK TEST	0.00	19:13

