

SOLUTION CHANGE

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 04/01/2020

OPERATOR NAME:  
 RICHARD J ARNOLD  
 Permit #: 33860  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:08
INTERNAL STANDARD	VERIFIED	19:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:09
BLANK TEST	0.00	19:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:11
BLANK TEST	0.00	19:12

