

SOLUTION CHANGE

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 12/02/2020

OPERATOR NAME:
 RICHARD J ARNOLD
 Permit #: 33860
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:52
BLANK TEST	0.00	00:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54

