

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 11/18/2020

OPERATOR NAME:
RICHARD J ARNOLD
Permit #: 33860
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:35
INTERNAL STANDARD	VERIFIED	05:35
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:35
BLANK TEST	0.00	05:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:36
BLANK TEST	0.00	05:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:38
BLANK TEST	0.00	05:39

