

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 01/09/2021

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:14
INTERNAL STANDARD	VERIFIED	11:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:14
BLANK TEST	0.00	11:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:16
BLANK TEST	0.00	11:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	11:17
BLANK TEST	0.00	11:18

