

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 01/17/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:58
INTERNAL STANDARD	VERIFIED	02:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:00
BLANK TEST	0.00	03:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:01
BLANK TEST	0.00	03:02

