

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 01/04/2021

OPERATOR NAME:  
CORBET W SULLIVAN  
Permit #: 45898  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:52
INTERNAL STANDARD	VERIFIED	06:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:53
BLANK TEST	0.00	06:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:54
BLANK TEST	0.00	06:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:55
BLANK TEST	0.00	06:56

CS #623

