

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 01/18/2021

OPERATOR NAME:  
CHRISTOPHER J MURRAY  
Permit #: 46226  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:55
INTERNAL STANDARD	VERIFIED	00:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:55
BLANK TEST	0.00	00:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:57
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:59

CM #821

