

SOLUTION CHANGE

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 01/01/2021

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54

CM 821

