

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 12/31/2020

OPERATOR NAME:
DAVID F BUXTON
Permit #: 33398
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 12:54 |
| INTERNAL STANDARD | VERIFIED | 12:54 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 12:54 |
| BLANK TEST | 0.00 | 12:55 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 12:55 |
| BLANK TEST | 0.00 | 12:56 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.10 | 12:57 |
| BLANK TEST | 0.00 | 12:58 |

Sgt David
[Signature]

