

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/11/2021

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:35
INTERNAL STANDARD	VERIFIED	06:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:36
BLANK TEST	0.00	06:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:37
BLANK TEST	0.00	06:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	06:38
BLANK TEST	0.00	06:39

