

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 01/18/2021

OPERATOR NAME:  
 CHRISTOPHER J DAVIDSEN  
 Permit #: 40077  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:22
BLANK TEST	0.00	07:23

