

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 01/04/2021

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:42
INTERNAL STANDARD	VERIFIED	12:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:42
BLANK TEST	0.00	12:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:43
BLANK TEST	0.00	12:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:45
BLANK TEST	0.00	12:45

