

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 01/11/2021

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:29
INTERNAL STANDARD	VERIFIED	08:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:31
BLANK TEST	0.00	08:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:33

