

Control Test Record		Simulator Information	
Serial Number:	113306	Lot:	20170
Test Date:	06/23/2020 15:22:19	Nominal:	0.100
Location:	03	Test Results	
GRANVILLE POLICE DEPT.		Alcohol 1:	0.09
Supervisor Information		Alcohol 2:	0.10
First Name:	Ryan 5	Alcohol 3:	0.10
Last Name:	Pedone	Solution Change: <input type="checkbox"/>	
Permit:	35878	Status	
Agency:	GRANVILLE VILLAGE PD	Status Code:	
Plot 1 of 3		Date Time:	
		Message:	