

SUPERVISOR TEST

CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706

Date: 01/25/2021

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:16
INTERNAL STANDARD	VERIFIED	07:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:17
BLANK TEST	0.00	07:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:18
BLANK TEST	0.00	07:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:19
BLANK TEST	0.00	07:20

