

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 10/13/2020

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:24
INTERNAL STANDARD	VERIFIED	08:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:24
BLANK TEST	0.00	08:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:25
BLANK TEST	0.00	08:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:26
BLANK TEST	0.00	08:27

