

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dnt: 113806

Date: 01/23/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:58
BLANK TEST	0.00	13:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:00
BLANK TEST	0.00	14:01

