

SUPERVISOR TEST

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 10/20/2020

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:18
INTERNAL STANDARD	VERIFIED	20:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:18
BLANK TEST	0.00	20:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:20
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:21
BLANK TEST	0.00	20:22

