

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 07/20/2020

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:58
INTERNAL STANDARD	VERIFIED	19:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	19:59
BLANK TEST	0.00	20:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:00
BLANK TEST	0.00	20:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:01
BLANK TEST	0.00	20:02

