

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 01/17/2021

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:04
INTERNAL STANDARD	VERIFIED	21:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:05
BLANK TEST	0.00	21:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:07
BLANK TEST	0.00	21:08

