

SOLUTION CHANGE

-----  
 FORT EDWARD  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 115708  
 -----

Date: 02/07/2021

OPERATOR NAME:  
 SGT DEAN E WATKINS  
 Permit #: 39523  
 Operator Agency/Dept: FORT EDWARD VILL. PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:58
INTERNAL STANDARD	VERIFIED	22:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:00
BLANK TEST	0.00	23:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:01
BLANK TEST	0.00	23:02

