

SOLUTION CHANGE

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 01/03/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:30
INTERNAL STANDARD	VERIFIED	21:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:31
BLANK TEST	0.00	21:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:34
BLANK TEST	0.00	21:34

